



# SinoGeneClon Biotech Co.,Ltd

## INTRODUCTION

### Rat Anti-Smooth Muscle Antibody(ASMA) ELISA Kit

FOR RESEARCH USE ONLY. Not for clinical diagnosis use

Catalog No : SG-20639

For the qualitative determination of Rat ASMA concentrations.

Reactivity:	Rat
Method Type:	Sandwich ELISA Detection
Quantity:	96 tests
Sample type:	serum, plasma, Urine,tissue homogenates, cell culture supernates

### Components:

Assay plate (12 × 8 coated Microwells)	1
Positive control	1×0.5ml
Negative control	1×0.5ml
HRP-Conjugate Reagent	1×6ml
Sample Diluent	1×6ml
Chromogen Solution A	1×6ml
Chromogen Solution B	1×6ml
Stop Solution	1×6ml
Wash Solution	1×20ml×30 fold
User manual	1
Adhesive Strip	2

### Product Principle:

The kit assay ASMA in the sample, use purified antigen to coat microtiter plate, make solid-phase antigen, then add ASMA to wells, Combined With ASMA ,after washing and removing non-combinative antibody and other components ,then Combined antigen which with HRP labeled become antigen - antibody - enzyme- antigen complex, after washing Completely, Add TMB substrate solution, TMB substrate becomes blue color At HRP enzyme-catalyzed, reaction is terminated by the addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. Compared with the CUTOFF value, according to this to judge ASMA exist in the sample or not

### Specimen requirements:

- 1.Serum-coagulation at room temperature for 10-20 min, centrifuge at the speed of 2000-3000 rpm for 20-min. Remove supernatant, if precipitation appeared, Centrifuge again. Assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.
- 2.Plasma-use suited EDTA or citrate plasma as an anticoagulant, centrifuge at the speed of 2000-3000 rpm for 20-min. Remove supernatant, if precipitation appeared, centrifuge again. Assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles. Centrifuge the sample again after thawing before the assay .
- 3.Urine-collect sue a sterile container, centrifuge at the speed of 2000-3000 rpm for 20-min. Remove supernatant, if precipitation appeared, Centrifuge again. The Operation of Hydrothorax and cerebrospinal fluid reference to it. Assay immediately or aliquot and store samples at -20°C or -80°C.
- 4.Cell culture supernatant-detect secretory components, Remove particulates by centrifugation for 20-min at the speed of 2000-3000 rpm. Remove supernatant detect the composition of cells, dilute cell suspension with PBS (PH7.2-7.4), Cell concentration reached 1 million / ml, repeated freeze-thaw cycles, damage cells and release of intracellular components, centrifugation 20-min at the speed of 2000-3000 rpm. remove supernatant, If precipitation appeared, Centrifugal again. Assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.



# SinoGeneClon Biotech Co.,Ltd

## INTRODUCTION

---

5. Tissue samples- After cutting samples, check the weight, Pipette PBS(PH7.2-7.4), Rapidly frozen with liquid nitrogen, maintain samples at 2-8°C after melting, Pipette PBS(PH7.4), homogenized by hand or Grinders, centrifugation 20-min at the speed of 2000-3000 rpm. Remove supernatant. Assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles. Centrifuge the sample again after thawing before the assay.

Note:

1. Extract as soon as possible after Samples collection, and should be tested as soon as possible after the extraction. If not, samples must be stored at -20°C ( $\leq 1$  month) or -80°C ( $\leq 2$  months) to avoid loss of bioactivity and contamination.
2. Can't detect the sample which contain NaN<sub>3</sub>, because NaN<sub>3</sub> inhibits HRP active.

### Reagent preparation

Wash Buffer (1x) - If crystals have formed in the concentrate, warm up to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 ml of Wash Buffer Concentrate (30 x) into deionized or distilled water to prepare 600 ml of Wash Buffer (1 x).

### Assay procedure:

Step 1: Number: determine the number of well to be used and store unused wells in 4 °C. Set a blank well without any solution

Step 2: Prepare sample: pipette Positive control and Negative control 50µl to the well respectively. Controls need test in duplicate.

Pipette sample Diluent 40µl and testing sample 10µl to testing sample well. Pipette sample to the bottom of well, don't touch the wall as far as possible, and mix gently.

Step 3: Incubate: Cover with the adhesive strip provided, incubate for 30 min at 37°C.

Step 4: Configure liquid: Dilute wash solution 30-fold with distilled water.

Step 5: Washing: Uncover the adhesive strip, discard liquid, pipette washing buffer to every well, still for 30s then drain, repeat 5 times.

Step 6: Add enzyme: Pipette HRP-Conjugate reagent 50µl to each well, except blank well.

Step 7: Incubate: Operation with 3

Step 8: Washing: Operation with 5.

Step 9: Color: Pipette Chromogen Solution A 50ul and Chromogen Solution B 50ul to each well, avoid the light preservation for 15 min at 37°C.

Step 10: Stop the reaction: Pipette Stop Solution 50µl to each well, stop the reaction (the blue change to yellow).

Step 11: Calculate: take blank well as zero. Read absorbance at 450nm after pipette Stop Solution within 15min.

### Calculation of result:

Test validity: the average of Positive control well  $\geq 1.00$ ; the average of Negative control well  $\leq 0.10$ .

Calculate Critical (CUT OFF): Critical = the average of Negative control well + 0.15.

Negative judgement: sample OD < Calculate Critical (CUT OFF) is ASMA Negative .

Positive judgement: sample OD  $\geq$  Calculate Critical (CUT OFF) is ASMA Positive .

### Expiration:

Twelve months [see label on the outer box for the specific date]

### Storage conditions:

The unopened kit shall be stored at [2-8 °C]

For opened kit can be stored at [2-8 °C] for up to 1 month. If not be used recently, the standard should be kept in -20 °C



# SinoGeneClon Biotech Co.,Ltd

## INTRODUCTION

---

### Attention:

- 1: The kit takes out from the refrigeration should be balanced 15-30 minutes in the room temperature, if the coated ELISA plates have not been used up after opening, the plate should be stored in sealed bag.
- 2: Washing buffer will Crystallization separation, it can be heated in water to dissolve.
- 3: Pipette sample with pipettors each step, and proofread its accuracy frequently to avoid the experimental error. Pipette sample within 5 min, if the number of sample is big, recommend using multichannel pipettor.
- 4: If the testing material concentration is excessively high (The sample OD is higher than the first standard well), please dilute the sample (n-fold).
- 5: Adhesive Strip only limits the disposable use to avoid cross-contamination.
- 6: The substrate should evade the light to be preserved.
- 7: Please refer to the user instruction strictly, the test result determination must take the microtiter plate reader as a standard.
- 8: The preparation of samples and all the reagents should refer to infective material process.
- 9: Do not mix reagents with those from other lots.

### Washing method:

Manually washing method: shake away the remained liquid in the enzyme plates; place some bibulous papers on the test-bed, and flap the plates on the upside down strongly. Inject at least 0.35ml after-dilution washing solution into the well, and marinate 1~2 minutes. Repeat this process according to your requirements.

Automatic washing method: if there is automatic washing machine, it should only be used in the test when you are quite familiar with its function and performance .

#### If You Have Problems

Technical Service Contact information

Email: [tech@sinogeneclon.com](mailto:tech@sinogeneclon.com)

Web: [www.sinogeneclon.com](http://www.sinogeneclon.com)

In order to obtain higher efficiency service, please ready to supply the lot number of the kit to us (found on the outside of the box).